

manual[®]**CL 119**

COATED LATEX

PRODUCT INFORMATION

Type: Powder-free, polymer coated and non-sterile examination glove
Primary Material: Natural Rubber Latex
Design and Feature: Ambidextrous, straight fingers, textured fingertips, beaded cuff
Colour: Off-white to yellow
Powder: ≤ 2 mg/glove (powder-free), no powder lubricant added
Packing: 100 gloves by weight per box, 10 boxes per shipping carton
Shelf Life: 5 years upon date of manufacture

Dimensions:

Gloves Size	Reorder #	Palm Width (mm)	Min. Length (mm)
Extra-small	CL119-00	< 80	240
Small	CL119-01	80 ± 10	240
Medium	CL119-02	95 ± 10	240
Large	CL119-03	110 ± 10	240
Extra-large	CL119-04	≥ 110	240

Thickness:

Location of Thickness Measurement	Single Wall (mm)	Typical Value
Finger (at 13 ± 3 mm from the fingertip)	Min. 0.08	0.12
Palm (at center of palm)	Min. 0.08	0.10

Physical Properties:

Parameters	Before Aging	After Aging
Tensile Strength (MPa)	Min. 18	Min. 14
Elongation (%)	Min. 650	Min. 500
Force At Break (N) – spec.	Min. 6.0	Min. 6.0
Typical Force At Break (N)	7.0	7.0

Pre-Shipment Inspection:

Inspection Criteria	Inspection Level	AQL
1000ml Water Leak Test	G-1	1.5

Product Conformance:

Certified as Personal Protective Equipment of Complex Design Category III as per PPE Regulation (EU) 2016/425, type tested in accordance with EN ISO 374-1:2016, EN ISO 374-5:2016 and EN ISO 21420:2020. Medical Device of Class I as per Medical Device Regulation (EU) 2017/745, tested in accordance with EN 455 parts 1-4.

