

DATA SHEET

DERMAGRIP®

Elbow Length Procedure Gloves

Powder Free Latex Examination Glove, sterile

Reorder#/Size		Small (6½) :	D3965-15
		Medium (7½) :	D3975-15
		Large (8½) :	D3985-15
		Extra Large (9) :	D3990-15
Design and Features		Hand specific, curved fingers, textured surface at palm and fingers and beaded cuff	
Type		Elbow length, powder free and gamma-sterilized examination glove	
Material		Natural rubber latex	
Colour		Natural	
Surface Treatment		Chlorination	
Protein Content (µg/g)		≤ 50	
Powder Residues (mg/glove)		≤ 2 (powder free)	
Dimensions – Palm width (mm)		Small (6½) :	80 ± 10
		Medium (7½) :	95 ± 10
		Large (8½) :	110 ± 10
		Extra Large (9) :	≤ 110
Dimensions - Length (mm)		Small (6½) :	min. 490 ± 10
		Medium (7½) :	min. 490 ± 10
		Large (8½) :	min. 490 ± 10
		Extra Large (9) :	min. 490 ± 10
Single wall thickness (mm)	Finger	Specification	Typical Value
	Palm	0.38 ± 0.06	0.36
	Cuff	0.34 ± 0.05	0.32
		0.24 ± 0.04	0.21
Physical Properties	Force at break (N) - specification	Before aging :	min. 9
	Typical Force at break (N)	After aging :	min. 9
Packing mode	Inner Wrapper	1 pair of gloves	
	Pouch	1 inner wrapper	
	Dispenser	25 pouches	
	Outer carton	4 dispensers	
Glove marking		No marking	
Lot number		Lot number structure: YMMPPPPSS Y = year of packing P = internal reference no. MM = month of packing SS = size	
Shelf life		5 years from date of manufacture	
Pre-shipment Quality Inspection	Dimensions	N=13 (EN 455-2)	
	Physical Properties	N=13 (EN 455-2)	
	1000ml Water Leak	G-I, AQL 1.5 (EN 455-1)	
	Visual Inspection	G-I, AQL 2.5	
Product Conformance		- Medical Device in compliance with European Medical Device Regulation 2017/745 (CE Class I-s): EN 455 Parts 1, 2, 3 & 4 - ASTM D3578	
Quality Assurance		- ISO 9001 - ISO 13485 Quality Management System - ISO 14001 Environmental Management System - US FDA Quality System Regulation (QSR)	

Note: The above information is a guideline of typical performance values and characteristics of the product and not to be used as actual product specifications.